

Lactic acid (LA) Content Assay Kit

Note: Take two or three different samples for prediction before test.

Operation Equipment: High performance liquid chromatography

Catalog Number: BC2234

Size: 50T/48S

Product Composition: Before use, please carefully check whether the volume of the reagent is consistent with the volume in the bottle. If you have any questions, please contact Solarbio staff in time.

Reagent name	Size	Preservation Condition
Extract solution	Liquid 80 mL×1	2-8°C
Reagent I	Liquid 2 mL×1	2-8°C
Reagent II	Powder×2	2-8°C
Standard	Powder×1	2-8°C

Solution Preparation:

1. Reagent II: Dissolve one bottle of Reagent II into 1000 mL of ultrapure water, add 0.9 mL of Reagent I, mix well, and obtain mobile phase A.
2. 1000 mL of prepared mobile phase A was suctioned with a membrane. (Prepared mobile phase A was suctioned with 0.22 μm aqueous membrane).
3. Ultrasonicate the filtered Mobile Phase A for 20 minutes to remove air bubbles.
4. Preparation of Standard: Before use, add 2 mL of distilled water to prepare a 10 mg/mL lactic acid standard solution, and store it at 2-8°C in a sealed container, avoiding direct sunlight. The 10 mg/mL lactic acid standard solution was diluted with distilled water to 2500 μg/mL, 1250 μg/mL, 625 μg/mL, 62.5 μg/mL, and 31.25 μg/mL lactic acid standard solutions, respectively (The prepared standard concentrations are for reference only and can be adjusted according to the actual sample concentration). Store (sealed) at 4°C in the dark, filter into the brown sample bottle with an aqueous syringe filter before testing, and wait for testing.

Product Description:

Lactic acid is an organic acid derived from cornstarch through biological fermentation and purification processes. It is a colorless, viscous liquid that exhibits acidic reactions in aqueous solutions. It is miscible with water, ethanol, or ether in any proportion and is insoluble in chloroform. Due to its levorotatory characteristic, it has excellent biocompatibility, can integrate with mammals, and can directly participate in human metabolism without any side effects. It is widely used in food, pharmaceutical, and other fields.

Lactic acid has an absorption peak at 210 nm, which allows for its content to be determined using high-performance liquid chromatography.

Reagents and Equipment Required but Not Provided:

High-performance liquid chromatography (HPLC) instrument (with Polaris C18-A column (4.6×250 mm) and variable wavelength detector (VWD), desk centrifuge, adjustable pipette, mortar/homogenizer, EP tubes (2 mL), needle filters (for organic and aqueous solutions), syringes, filtration apparatus, filter membranes (for aqueous and organic solutions), 50 brown sample vials (1.5 mL), and ultrapure water.

Operation procedure

I. Extraction of lactic acid:

According to the ratio of mass (g) to extraction solution volume (mL) of 1:5~10, it is recommended to weigh 0.15 g of fresh sample, grind it thoroughly, add 1 mL of extraction solution, seal it, mix well, and place it in a 75°C water bath for 20 minutes of extraction. Centrifuge at 12000 rpm for 10 minutes, take the supernatant, add 0.5 mL of extraction solution to the filter residue again, shake to mix well, and place it in a 75°C water bath for 20 minutes of extraction. Mix the two extractions, centrifuge at 12000 rpm for 10 minutes, take the supernatant, and filter it into a brown sample vial using an aqueous needle filter before testing (if the supernatant is too dark or concentrated, it can be diluted and filtered again before testing).

II. Determination procedure:

1. Turn on the computer, switch on all modules of the liquid chromatograph, install the chromatographic column, open the software, set the injection volume to 10 µL in the method group, set the column temperature to 30°C, the flow rate to 0.4 mL/min, the wavelength to 210 nm, and the run time to 25 minutes. Save the method group after setting.
2. Clean the column with the corresponding mobile phase, balance the column with mobile phase A, and start adding samples after the baseline stabilizes.
3. Detect the standard solution to be tested with an injection volume of 10 µL. Lactic acid can be separated within 25 minutes, and the retention time of lactic acid is around 12.3 minutes (the retention time may vary depending on factors such as the system, column, pH of the mobile phase, and temperature, and is only provided as a reference).
4. Detect the sample solution to be tested with an injection volume of 10 µL, and measure the peak area of lactic acid at the corresponding retention time.

III. Calculation:

Plot a standard curve for lactic acid with the concentration of the standard solution (µg/mL) as the abscissa and the peak area as the ordinate. Substitute the peak area of the sample into the standard curve to calculate the concentration x (µg/mL) of lactic acid in the extraction solution.

The content of lactic acid (µg/g) = $x \times V \div W = 1.5x \div W$

V: The volume of the extract solution, 1.5 mL;

W: Sample weight(g).

For samples tested after dilution, multiply by the corresponding dilution factor before calculation.

Note:

1. After the test is completed, the chromatographic column should be rinsed with a high concentration of ultrapure water phase (about 20-30 column volumes) to prevent blockage, followed by rinsing with a high concentration of organic phase. Finally, rinse the column according to its specific type to prevent damage.
2. The dilution factor of the standard solution should be determined based on the concentration of lactic acid in the sample. The peak area of lactic acid in the sample must fall within the range of peak areas obtained from standard solutions of different concentrations. The suggested dilution factor for the standard solution is only a reference. If the concentration of lactic acid in the sample is too high, it is recommended to dilute it before measurement.
3. If the sample volume is too large, it is advisable to test a standard solution (just one) once a day to confirm the corresponding retention time. All solutions to be tested should be brought to room temperature before testing.