

Beijing Solarbio Science & Technology Co.,Ltd. One-stop solution for life science research.

# Plant Tissue Fructose Content Assay Kit

Note: It is necessary to predict 2-3 large difference samples before the formal determination.

**Operation Equipment:** Spectrophotometer

**Catalog Number:** BC2450

Size:50T/24S

## **Components:**

Name of reagent	Size	storage	
<b>Extract Solution</b>	Liquid 30 mL × 1 bottle	2-8 °C	
Reagent I	Liquid 2.5 mL $\times$ 1 bottle	2-8 °C	
Reagent II	Liquid 45 mL × 1 bottle	2-8 °C	
Reagent III	Liquid 15 mL × 1 bottle	2-8 °C	
Reagent IV	Powder $\times$ 1 bottle	RT S	
Standard	Powder × 1	2-8 °C	

Preparation of solution:

- 1. Standard: Add 0.695mL distilled water to dissolve before use to obtain 80 mmol/L fructose standard, and store it at 2-8 °C for 4 weeks;
- Preparation of 8mmol/L standard: take 100µL 80mmol/L standard, add 900µL distilled water, fully mix. prepare 8mmol/L standard for standard tube determination

# **Product Description:**

Fructose is one kind of ketohexose and an isomeride of glucose, which exists widely in the  $p\mu lp$  of fruits and honey with free state. It can be combined with glucose to form sucrose. Fructose is the sweetest monosaccharide and is used widely in the production of food, medicine and health care products.

In acidic conditions, fructose reacts with resorcinol to form colored substances which has characteristic absorption peaks at 480 nm.

### **Reagents and Equipments Required but Not Provided:**

Spectrophotometer, water bath, centrifuge, adjustable transferpettor, 1 mL glass cuvette, mortar/homogenizer, distilled water.

### **Operation procedure:**

### **I. Sample Preparation**

- (1) Weigh about 0.05g sample and grind it at room temperature; Add 0.5mL of extract Solution, grind it properly and transfer it to the covered centrifuge tube quickly;
- (2) Place it in a water bath at 80 °C for 10min (wrap sealing film to prevent lid explosion), shake it for 3-5 times, cool it to room temperature, centrifuge it at 4000g at room temperature for 10min, and take 0.4mL of supernatant;



B2450 -- Page 1 / 3

For research use only. Do not use for clinical, diagnostic, food, cosmetic testing and other purposes.



- (3) Add a small amount (about 2mg) of reagent 4 and decolor at 80 °C for 30min (wrap sealing film to prevent cap explosion);
- (4) Then add 0.4mL of extract Solution, 4000g, centrifuge at RT for 10min, and take the supernatant for determination.

**Note:** If there is any volume change, the supernatant in step (2) and the extract in step (4) shall be added according to the volume ratio of 1:1;

(2) To prevent explosion cover, you can use the spiral mouth EP tube or with a lock buckle EP tube (wrapped sealing film), you can also tied 1 hole on the cover of ordinary EP tube (wrapped sealing film)

# II. Detection

1. Preheat spectrophotometer for 30 min, adjust the wavelength to 480 nm and set the counter to zero with distilled water.

		15.7		
Reagent name (µL)	Test tube (A <sub>T</sub> )	Control tube (A <sub>C</sub> )	Standard tube (As)	Blank tube (A <sub>B</sub> )
Sample	120	120	-	-
Standard		-	120	-
Distilled water	80	-	80	200
Reagent I	-	80	- Une	
COLSTON	-	Water bath at 100 °C for 10min	-	Solarvice
Reagent II	760	760	760	760
Reagent III	240	240	240	240

2. Add these reagent to a 1.5 mL EP tube:

Vortex and mix evenly, place in a 100 °C water bath for accurate reaction for 10min (wrap sealing film to prevent cover explosion), cool to room temperature, take the reaction solution into a 1mL glass cuvette to measure the absorbance value at 480nm, record as  $A_T$ ,  $A_C$ ,  $A_S$ ,  $A_B$ , and calculate  $\Delta A = A_T - A_C$ ,  $\Delta A_S = A_S - A_B$ . Standard tube and blank tube only need to measure 1-2 times.

# **Calculation:**

1. Protein concentration:

Fructose (mg/mg prot) =  $C \times \Delta A \div \Delta As \times V_S \div (Cpr \times V_S) \times 180 \times F = 1440 \times \Delta A \div \Delta As \div Cpr \times F$ 

2. Sample weight:

Fructose (mg/g fresh weight) =  $C \times \Delta At \div \Delta As \times Ve \div W \times F = 1.44 \div \Delta A \div \Delta As \div W \times F$ 

C: Standard concentration, 8mmol/L;

Cpr: Sample concentration (mg/mL);

W: Sample weight(g);

Vs: sample volume, 0.12mL;

Ve: Extraction volume,  $1 \times 10^{-3}$  L;

B2450 -- Page 2 / 3

For research use only. Do not use for clinical, diagnostic, food, cosmetic testing and other purposes.



180: Fructose Relative molecular mass, 180mg/mmol.

# Note:

1. If the  $\Delta A$  is greater than 1.3, it is recommended that the sample be diluted with water for determination. Pay attention to multiplying the dilution multiple in the calculation formula; if the  $\Delta A$  is less than 0.02, it is recommended to increase the sample mass and re-extract.

2. Boiling or high temperature process can use spiral mouth EP tube or with lock buckle EP tube and wrap sealing film; also can be wrapped in the film of ordinary EP tube cover 1 holes to prevent explosion.

3. The protein has been removed during the pretreatment process. If the protein concentration is used to calculate, another tissue shall be taken for re-extraction.

# **Experimental examples:**

1. Take 0.1023g of mango fruit, carry out pretreatment according to the sample treatment steps, take the supernatant and dilute it 4 times according to the determination steps, measure and calculate  $\Delta A = 0.439-0.207=0.232, \Delta A_s = 0.62-0.002=0.618$ , and calculate according to the sample mass:

Fructose content (mg/g mass) =  $1.44 \times \Delta A \div \Delta A_S \div W \times F = 21.137$  mg/g mass.

2. Take 0.1032g carrot, carry out pretreatment according to the sample treatment steps, take supernatant, dilute 4 times according to the determination steps, measure with 1mL glass cuvette and calculate  $\Delta A = 0.689-0.358=0.331$ ,  $\Delta A_S = 0.62-0.002=0.618$ , and calculate according to the sample mass:

Fructose content (mg/g mass) =  $1.44 \times \Delta A \div \Delta A_{s} \div W \times F = 29.894$  mg/g mass.

### **References:**

[1] Varandas S, Teixeira M J, Marques J C, et al. Glucose and fructose levels on grape skin: interference in Lobesia botrana behaviour[J]. Analytica chimica acta, 2004, 513(1): 351-355.

### **Related Products:**

BC0340/BC0345 Glycogen Assay Kit BC2460/BC2465 Plant Sucrose Content Detection Kit BC2500/BC2505 Glucose Assay Kit BC0330/BC0335 Trehalose Content Assay Kit



B2450 -- Page 3 / 3

For research use only. Do not use for clinical, diagnostic, food, cosmetic testing and other purposes.