

Sorbitol Content Assay Kit

Note: Take two or three different samples for prediction before test.

Operation Equipment: Spectrophotometer

Catalog Number: BC2520

Size: 50T/48S

Components:

Reagent I: 15 mL×1, storage at 2-8°C.

Reagent II: 15 mL×1, storage at 2-8°C.

Standard: 10 mg×1, storage at 2-8°C. Dissolve the standard with 1 mL of distilled water at a concentration of 10 mg/mL. The unused reagent can be stored at 4°C for 2 weeks.

Product Description

Sorbitol is widely found in animals, plants, microorganisms and cultured cells. It's a form of sugar transport, and also closely related to biological resistance and food flavor. Therefore, it is often necessary to detect the changes of sorbitol content in sugar metabolism, stress resistance and food research.

Sorbitol can form blue complex with copper ions in alkaline solution, which has characteristic absorption peaks at 655 nm.

Reagents and Equipment Required but Not Provided.

Spectrophotometer, desk centrifuge, transferpettor, 1 mL glass cuvette, mortar/homogenizer, ice and distilled water.

Procedure:

I. Sample extraction:

Suggest that weight about 0.2 g tissue with 2 mL distilled water. Fully grinding, 95°C water bath for 10 minutes (cover tightly to prevent moisture loss), after cooling, 8000 g 25°C centrifuge for 10 min. Supernatant is used for test.

II. Determination procedure:

1. Preheat the spectrophotometer 30 min, adjust wavelength to 655 nm, set zero with distilled water.
2. Preparation of standard solution: before use, dilute 10mg/mL standard solution with distilled water to 4, 2, 1, 0.5, 0.25, 0.125, 0 (blank tube) mg/mL standard solution to be measured.
3. Sampling table (add the following reagents in EP tube)

Reagent (μL)	Standard Tube	Test Tube
Reagent I	150	150
Reagent II	140	140
Sample	-	1000
Standard	1000	-

Mix well, place at room temperature for 15 min, centrifuge at 8000 g and 25°C for 10 min and

take

supernatant to detect the absorbance at 655 nm. $\Delta A_t = A_t - A_b$, $\Delta A_s = A_s - A_b$. The standard curve and blank tubes should only be measured 1-2 times.

III. Calculation:

1. A standard curve was created from the concentration of the standard tube (y, mg/mL) and the absorbance A (x, ΔA_s). Based on the standard curve, the sample concentration (y, mg/mL) was calculated by bringing the A_t (x, ΔA_t) into the equation.

2. Sample weight

$$\text{Sorbitol (mg/g fresh weight)} = y \times V_1 \div (W \times V_1 \div V_2) = 2 \times y \div W$$

3. Protein concentration

$$\text{Sorbitol (mg/mg prot)} = y \times V_1 \div (V_1 \times C_{pr}) = y \div C_{pr}$$

V1: sample volume, 1 mL;

V2: extraction volume, 2 mL;

Cpr: sample protein concentration, mg/mL;

W: Sample weight, g.

Note:

If the measured absorbance value exceeds the linear range, the sample size can be increased or the sample can be diluted before measurement.

References:

Salvucci M E, Stecher D S, Henneberry T J. Heat shock proteins in whiteflies, an insect that accumulates sorbitol in response to heat stress[J]. Journal of Thermal Biology, 2000, 25(5): 363-371.

Related Products:

BC0340/BC0345	Glucogen Content Assay Kit
BC2540/BC2545	Cellulase(CL) Activity Assay Kit
BC0330/BC0335	Trehalose Content Assay Kit
BC2490/BC2495	Blood Glucose Content Assay Kit
BC2530/BC2535	Sorbitol Dehydrogenase(SDH) Activity Assay Kit

Technical Specification:

The detection limit: 0.0261 mg/mL

Linear range: 0.0625-4 mg/mL