

Beijing Solarbio Science & Technology Co.,Ltd. One-stop solution for life science research.

Ornithine aminotransferase (δ-OAT) Activity Assay Kit

Note: Take two or three different samples for prediction before test.

Detection equipment: Spectrophotometer/microplate reader

Cat No: BC4425

Size: 100T/96S

Components:

Extract solution: 110 mL×1, store at 4°C and protect from light.

Reagent I: 40 mL×1, store at 4°C.

Reagent II: Powder×2, store at 4°C and protect from light. Add 5 mL of Reagent I when the solution will be used. Unused reagent can store at -20°C after packing.

Reagent III: Powder×2, store at 4°C and protect from light. Add 5 mL of Reagent I when the solution will be used. Unused reagent can store at -20°C after packing.

Reagent IV: Powder×1, store at -20°C and protect from light. Add 10 mL of distilled water when the solution will be used. Unused reagent can store at -20°C for one week after packing.

Description:

Ornithine aminotransferase (δ -OAT) is one of the key enzymes for the synthesis of proline by ornithine as a precursor, which plays an important role in adapting plants to stress. Ornithine and α -ketoglutarate can undergo acyl transfer reaction under the action of δ -OAT and NADH to produce NAD and pyrrolaldehyde-5-carboxylic acid (P5C). NADH has a special absorption peak at 340 nm. By measuring the change in absorbance at 340 nm, the level of δ -OAT activity can be calculated.

Required but not provided

Spectrophotometer, low temperature centrifuge, water-bath, transferpettor, micro quartz cuvette/96 UV well plate, homogenizer, ice, distilled water and EP tubes.

Protocol:

I. Crude enzyme extraction:

1. Tissue:

The mass of tissue (g): the volume of extract (mL) is $1:5\sim10$ (it is suggested to take about 0.1 g of tissue, add 1 mL of Extract solution), fully grinding on ice. Centrifuge at 12000 rpm for 10 minutes at 4°C, take the supernatant and place it on ice for test.

2. Bacteria or cells

Collecting bacteria or cells into the centrifuge tube. The supernatant is discarded after centrifugation. it is suggested to take about 5 million bacteria/cell and add 1 mL of extract solution. Bacteria or cell is splitted by ultrasonication(Power: 300 W, work time 3s, interval 7s, total time: 3 minutes). Centrifuge at 12000 rpm for 10 minutes at 4°C, take the supernatant and place it on ice for test. 3. Liquid samples: direct measurement.



II. Procedure

1. Preheat spectrophotometer for 30 minutes, adjust wavelength to 340 nm, set zero with distilled water.

2. Preheat Reagent II, Reagent III and Reagent IV at 37°C for 10 minutes.(Preheat as much reagent as needed).

3. Procedure test

Reagent (µL)	Blank tube (B)	Test tube (T)
Reagent II	60	60
Reagent III	60	60
Reagent IV	60	60
Sample	20	- 185
Distilled water	© -	20

Add reagents to micro quartz cuvette or 96 UV well plate orderly, mix thoroughly. Detect the absorbance at 340 nm at the time of 10 seconds record as A1. Then place dishes with the reaction solution in a 37 °C water bath or incubator for 10 minutes. Take it out and wipe it clean, immediately measure the absorbance at the time of 610 seconds which record as

 $A2. \Delta A_{T} = A_{T1} - A_{T2}, \Delta A_{B} = A_{B1} - A_{B2}, \Delta A = \Delta A_{T} - \Delta A_{B},$

Note: The Blank tube only needs to be measured one or twice.

III. Calculations of δ -OAT activity:

1. Protein concentration:

Unit definition: One unit of δ - OAT activity is defined as the amount of enzyme that per milligram of protein oxidation 1 mmoL of NADH per minute in the reaction system.

δ - OAT (U/minute prot)= $\Delta A \times V_{RV} \div (\epsilon \times d) \times 10^9 \div (V_S \div Cpr) \div T = 160.77 \times \Delta A \div Cpr \div d$

2. Sample weight:

Unit definition: One unit of δ - OAT activity is defined as the amount of enzyme that per gram of tissue oxidation 1 mmoL of NADH per minute in the reaction system.

 δ - OAT (U/minute fresh weight)= $\Delta A \times V_{RV}$ ÷(ε×d)×10⁹÷(W×V_S ÷V_E) ÷T= 160.77× ΔA ÷Wr÷d 3. Cell amount:

Unit definition: One unit of δ - OAT activity is defined as the amount of enzyme that per 10 thousand germ or cells oxidation 1 mmoL of NADH per minute in the reaction system.

δ - OAT (U/minute fresh weight)= $\Delta A \times V_{RV} \div (\varepsilon \times d) \times 10^9 \div (V_S \div N \div V_E) \div T = 160.77 \times \Delta A \div N \div d$

4. Liquid volume:

Unit definition: One unit of δ - OAT activity is defined as the amount of enzyme that per milliliter of liquid oxidation 1 mmoL of NADH per minute in the reaction system.

 δ - OAT (U/mL) = $\Delta A \div (\epsilon \times d) \times V_{RV} \times 10^9 \div V_S \div T = 160.77 \times \Delta A \div d$

V_{RV}: Total reaction volume, 0.01 L;

ε: Molar extinction coefficient, 6.22×103L/mol/cm;

d: Micro quartz cuvette light path, 1 cm; 96-well UV plate light path, 0.6 cm;



Vs: Sample volume, 0.02 mL; VE: Extract solution volume, 1 mL;

V_{RV}: Reaction volume, 2×10^{-4} L;

T: Reaction time(min), 10 minutes;

Cpr: Sample protein concentration, mg/mL;

- N: Total number of bacteria/cells, 10 million as a unit;
- W: Sample weight, g.

Note:

1. If $\Delta A > 0.5$, please dilute the sample to appropriate concentration, multiply dilute times in the formular. If ΔA is too small, increase the sample volume or prolong the enzymatic reaction time. 2. After adding the reagents in turn, mixed as quickly as possible and measured the OD, to reduce the error time.

3. ΔA_B generally does not exceed 0.05.

Experimental Examples:

Take 0.1g red bean stalks, carry out sample processing, and measure in the 96 well plate according to the operation steps. The calculation is: ΔAt=At1-At2=0.1058, ΔAb=Ab1-Ab2=0, ΔA=ΔAt -ΔAb = 0.1058, calculate the enzyme activity according to the sample weight:
δ-OAT (U/g weight) =160.77×ΔA÷W÷d=160.77×0.1058÷0.1÷1=170.09 U/g weight

Related Products:

BC1550/BC1555	Glutamic-pyruvic Transaminase(GPT) Activity Assay Kit
BC1560/BC1565	Glutamic-oxalacetic Transaminase(GOT) Activity Assay Kit
BC4160/BC4165	Proline Dehydrogenase(ProDH) Activity Assay Kit
BC4140/BC4145	Leucine Arylamidase(LAP) Activity Assay Kit

