

Albumin Content Assay Kit (Bromocresol Purple Colorimetry)

Note: Take two or three different samples for prediction before test.

Operation Equipment: Spectrophotometer

Catalog Number: BC5820

Size: 50T/48S

Components:

Extract Solution: 60 mL×1.

Reagent I: 0.24 mL×1, stored at 4°C.

Reagent II: 60mL×1, stored at 4°C.

Reagent III: 0.8mL×1, stored at 4°C.

Standard: 1 mL×1, 10 mg/mL, stored at -20°C. Before use, 50 μL 10 mg/mL albumin standard solution is taken and 150μL Extract Solution is added to prepare 2.5 mg/mL albumin standard solution.

Preparation of Color developing solution: According to the number of samples before use, the color developing solution is prepared according to the ratio of reagent I : reagent II : reagent III = 10μL : 3990μL : 40μL (4040μL, 4T).

Product Description

Albumin is the most important protein in human plasma. It is synthesized by the liver and is an important nutrient in the human body. It can maintain plasma osmotic pressure and can be combined with a variety of nutrients, hormones and drugs. Albumin content can reflect the nutritional status of the body, and can also check diseases that affect liver metabolic function, such as cirrhosis, liver injury, malnutrition, and malignant tumors.

In the acidic environment, the albumin molecule is positively charged, and combines with the negatively charged Bromocresol Purple (BCP) to form a green complex, which has a specific absorption peak at 603 nm. The absorbance of the complex is proportional to the concentration of albumin.

Reagents and Equipment Required but Not Provided.

Centrifuge, spectrophotometer, bowl/homogenizer/cell ultrasonic breaker, 1mL glass cuvette, ice and distilled water.

Procedure

I. Extraction of soluble protein in the sample:

1. Bacteria or cells: collecting bacteria or cells into the centrifuge tube, after centrifugation discard supernatant. According to bacteria or cells (10^4): Extract solution (mL) is 500~1000:1 to extract. It is suggested to add 1 mL of Extract solution to 5 million of bacteria or cells. Use ultrasonication to splitting bacteria and cells (placed on ice, ultrasonic power 200W, working time 3 seconds, interval 7 seconds,

repeat for 30 times). Centrifuge at 8000 ×g for 10 minutes at 4°C to remove insoluble materials and take the supernatant on ice before testing.

2. Tissue: according to tissue weight (g): Extract solution (mL) is 1:5~10 to extract. Add 1 mL of Extract solution to 0.1 g of tissue, and fully homogenized on ice bath. Centrifuge at 8000×g for 10 minutes at 4°C to remove insoluble materials, and take the supernatant on ice before testing.

3. Liquid samples: direct determination. If the liquid is turbid, centrifuge and take the supernatant for determination.

II. Measurement steps

a. Preheat the spectrophotometer for 30 minutes, adjust the wavelength to 603 nm and adjust zero with distilled water.

b. Operation table: (Add the following reagents to 1 mL glass cuvette)

Reagent Name (μL)	Test Tube (A _T)	Standard Tube (A _S)	Blank Tube (A _B)
Sample	100	-	-
Standard Solution	-	100	-
Extract Solution	-	-	100
Color developing solution	1000	1000	1000

Mix well, react at room temperature for 1 min, and measure the absorbance of each tube at 603 nm, which are recorded as A_T, A_S and A_B, respectively. Calculate $\Delta A_T = A_T - A_B$, $\Delta A_S = A_S - A_B$. Blank tube and standard tube only need to measure 1-2 times.

Note: The length of reacting time will affect the test results. It is recommended to react directly in 1 mL glass cuvette for 1 min to determine the absorbance value.

III. Calculation of Albumin Content:

1. Calculated according to Protein concentration:

$$\text{Albumin Content (mg/mg prot)} = \Delta A_T \times (C_S \div \Delta A_S) \times V_S \div (V_S \times C_{pr}) = 2.5 \times \Delta A_T \div \Delta A_S \div C_{pr}$$

2. Calculated according to the weight of the sample:

$$\text{Albumin Content (mg/mg weight)} = \Delta A_T \times (C_S \div \Delta A_S) \times V_S \div (W \times V_S \div V_E) = 2.5 \times \Delta A_T \div \Delta A_S \div W$$

3. Calculated according to liquid volume:

$$\text{Albumin Content (mg/mL)} = \Delta A_T \times (C_S \div \Delta A_S) \times V_S \div V_S = 2.5 \times \Delta A_T \div \Delta A_S$$

4. Calculated according to cell count:

$$\text{Albumin Content (mg/10}^6 \text{ cell)} = \Delta A_T \times (C_S \div \Delta A_S) \times V_S \div (N \times V_S \div V_E) = 2.5 \times \Delta A_T \div \Delta A_S \div N$$

C_S: Concentration of standard solution, 2.5mg/mL;

V_S: Sample volume, 0.1mL;

V_E: Extract solution volume, 1mL;

C_{pr}: Sample protein concentration, mg/mL;

W: Sample weight, g;

N: Total number of bacteria / cells, as 10⁶ a unit.

Notes:

1. If the $\Delta A_T < 0.005$ or A_T is close to the blank tube, the sample size can be increased before the determination ; if the determination of $\Delta A_T > 0.5$, it is recommended that the sample supernatant be diluted with the extract before determination. Note the simultaneous modification of the calculation formula.
2. If the sample is turbid after adding the chromogenic agent, it is recommended that the sample supernatant be diluted with the extract and then determined. Note the simultaneous modification of the calculation formula.

References:

- [1] Tada S, Yasukawa K, Yatomi Y, et al. A simple colorimetric assay to determine the concentration and proportion of human mercaptalbumin[J]. Practical Laboratory Medicine, 2022, 31: e00281.
- [2] Muramoto Y , Matsushita M , Irino T .Reduction of reaction differences between human mercaptalbumin and human nonmercaptalbumin measured by the bromcresol purple method[J].Clinica Chimica Acta, 1999, 289(1-2):69-78.